



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-05-21 14:44:50
Date of Next Testing: 2029-05-20 14:44:50
Product Name: **pH Buffers Solutions(pH8.00±0.01)**
CAS Number:
Specifications & Purity: Buffer Solution pH 8.00±0.01
Lot #: E2121016
SKU # **P196907**
Version **0**
Analysis #: **516022**

Parameter	Limit Values	Result
PH	7.99-8.01	8.00
Appearance (P196907)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager