



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-08-21 15:11:03
Date of Next Testing: 2027-08-20 15:11:03
Product Name: **pH Buffers Solutions(pH8.00±0.01)**
CAS Number:
Specifications & Purity: Buffer Solution pH 8.00±0.01
Lot #: H2318341
SKU # **P196907**
Analysis: **267173**

Parameter	Limit Values	Result
PH	7.99-8.01	8.01
Appearance (P196907)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager