

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-12-29 15:11:30

Date of Next Testing: 2026-12-28 15:11:30

Product Name: pH Buffer Standards DIN 19266 values (pH 4.005)

CAS Number:

Specifications & Purity: Buffer Solution pH 4.005

Lot #: L2229002

SKU #: P196923

Version: 0

Analysis #: 172466

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	3.995-4.015	4.00100
Appearance (P196923)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager