

Certificate of Analysis

Date of Testing: 2025-01-06 08:53:22

Date of Next Testing: 2029-01-05 08:53:22

Product Name: pH Buffer Standards DIN 19266 values (pH 6.865)

CAS Number:

Specifications & Purity: Buffer Solution pH 6.865

Lot #: A2506001

SKU # P196924

Version 0

Analysis #: 456895

Parameter	Limit Values	Result
PH(25 °C)	6.845-6.885	6.87
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager