

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-01-23 15:48:21
Date of Next Testing: 2030-01-22 15:48:21
Product Name: **pH Buffer Standards DIN 19266 values (pH 6.865)**
CAS Number:
Specifications & Purity: Buffer Solution pH 6.865
Lot #: A2619566
SKU #: **P196924**
Version: **0**
Analysis #: **638009**
Storage Temperature: Room temperature

| Parameter | Limit Values | Result |
|----------------------|------------------------|------------------------|
| pH(25 °C) | 6.845-6.885 | 6.86400 |
| Appearance (P196924) | Colorless clear liquid | Colorless clear liquid |



John Su
QA & QC Manager