

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-03-28 12:33:23
Date of Next Testing: 2027-03-27 12:33:23
Product Name: **pH Buffer Standards DIN 19266 values (pH 6.865)**
CAS Number:
Specifications & Purity: Buffer Solution pH 6.865
Lot #: C2327064
SKU #: **P196924**
Analysis: **204100**

Parameter	Limit Values	Result
PH(25 °C)	6.845-6.885	6.86
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager