

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-18 16:00:36
Date of Next Testing: 2029-07-17 16:00:36
Product Name: pH Buffer Standards DIN 19266 values (pH 6.865)
CAS Number:
Specifications & Purity: Buffer Solution pH 6.865
Lot #: G2517455
SKU #: P196924
Version: 0
Analysis #: 543611
Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	6.845-6.885	6.86500
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager