

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2023-09-14 18:46:04  
Date of Next Testing: 2027-09-13 18:46:04  
Product Name: **pH Buffer Standards DIN 19266 values (pH 6.865)**  
CAS Number:  
Specifications & Purity: Buffer Solution pH 6.865  
Lot #: I2306001  
SKU #: **P196924**  
Analysis: **275680**

Parameter	Limit Values	Result
PH(25 °C)	6.845-6.885	6.87
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid



John Su  
QA & QC Manager