

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2024-10-22 12:58:49  
Date of Next Testing: 2028-10-21 12:58:49  
Product Name: pH Buffer Standards DIN 19266 values (pH 6.865)  
CAS Number:  
Specifications & Purity: Buffer Solution pH 6.865  
Lot #: J2421909  
SKU #: P196924  
Version: 0  
Analysis #: 424745

Parameter	Limit Values	Result
PH(25 °C)	6.845-6.885	6.87
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su  
QA & QC Manager