

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing:	2025-11-14 14:10:49
Date of Next Testing:	2029-11-13 14:10:49
Product Name:	pH Buffer Standards DIN 19266 values (pH 6.865)
CAS Number:	
Specifications & Purity:	Buffer Solution pH 6.865
Lot #:	K2513427
SKU #	P196924
Version	0
Analysis #:	607140
Storage Temperature:	Room temperature

Parameter	Limit Values	Result
pH(25 °C)	6.845-6.885	6.87700
Appearance (P196924)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager