

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing:	2025-06-11 14:42:33
Date of Next Testing:	2029-06-10 14:42:33
Product Name:	pH Buffer Standards DIN 19266 values (pH 7.413)
CAS Number:	
Specifications & Purity:	Buffer Solution pH 7.413
Lot #:	G2112023
SKU #	P196925
Version	0
Analysis #:	527241
Storage Temperature:	Room temperature

Parameter	Limit Values	Result
pH(25 °C)	7.393-7.433	7.41300
Appearance (P196925)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager