

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-01-22 10:07:09

Date of Next Testing: 2027-01-22 10:07:09

Product Name: **Buffer, Reference Standard, pH 2.70 ± 0.01 at 25°C**

CAS Number:

Specifications & Purity: pH 2.70 ± 0.01(25°C)

Lot #: F2128064

SKU #: **P298591**

Version: **0**

Analysis #: **466018**

Parameter	Limit Values	Result
PH(25 °C)	2.69-2.71	2.70
Appearance (P298591)	Clear Colorless Liquid	Clear Colorless Liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager