

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-06-16 09:54:45

Date of Next Testing: 2027-06-16 09:54:45

Product Name: **Buffer, Reference Standard, pH 4.45 ± 0.01 at 25 °C**

CAS Number:

Specifications & Purity: pH 4.45 ± 0.01(25°C)

Lot #: H2127001

SKU # **P298592**

Version **0**

Analysis #: **529082**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	4.44-4.46	4.45000
Appearance (P298592)	Clear Colorless Liquid	Clear Colorless Liquid



John Su
QA & QC Manager