

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-11-27 09:07:34

Date of Next Testing: 2027-11-27 09:07:34

Product Name: **Buffer, Reference Standard, pH 1.68 ± 0.01 at 25°C**

CAS Number:

Specifications & Purity: pH 1.68 ± 0.01(25°C)

Lot #: K2513682

SKU #: **P298597**

Version: **0**

Analysis #: **613635**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	1.67-1.69	1.68500
Appearance (P298597)	Clear Colorless Liquid	Clear Colorless Liquid



John Su
QA & QC Manager