

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-10-19 11:39:28
Date of Next Testing: 2024-10-18 11:39:28
Product Name: **Buffer, Reference Standard, pH 7.38 ± 0.01 at 25°C**
CAS Number:
Specifications & Purity: pH 7.38 ± 0.01(25°C)
Lot #: A1921079
SKU #: **P298598**
Analysis: **145728**

Parameter	Limit Values	Result
PH(25 °C)	7.37-7.39	7.38
Appearance (P298598)	Clear Colorless Liquid	Clear Colorless Liquid



John Su
QA & QC Manager