

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-07-26 13:34:40

Date of Next Testing: 2025-07-25 13:34:40

Product Name: **Buffer, Reference Standard, pH 12.45 ± 0.01 at 25°C**

CAS Number:

Specifications & Purity: pH 12.45 ± 0.01(25°C)

Lot #: G2313007

SKU #: **P298605**

Version: **0**

Analysis #: **256058**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	12.44-12.46	12.45700
Appearance (P298605)	Clear Colorless Liquid	Clear Colorless Liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager