

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-03-12 14:18:46

Date of Next Testing: 2027-03-12 14:18:46

Product Name: **Reducing Sugar Content Assay Kit (Fehling's Titration Method)**

CAS Number:

Specifications & Purity: BioReagent

Lot #: ZJ26F0332963

SKU #: **R151034-CG40T**

Version: **1**

Analysis #: 030336

Parameter	Limit Value	Result
Seal Integrity Testing	Conform	Conform
Appearance(R151034A)	Colorless and transparent solution, free of sediment	Colorless and transparent solution, free of sediment
Odor(R151034A)	Odorless	Odorless
Density(R151034A)(0.96 - 1.04 g/mL)	Conform	Conform
Appearance(R151034B)	Blue and transparent solution, free of sediment	Blue and transparent solution, free of sediment
Density(R151034B)(0.98 - 1.05 g/mL)	Conform	Conform
Appearance(R151034C)	Colorless/light-yellow and transparent solution, free of sediment	Colorless/light-yellow and transparent solution, free of sediment
Odor(R151034C)	Odorless	Odorless
Density(R151034C)(1.07 - 1.14 g/mL)	Conform	Conform
Appearance(R151034D)	Colorless and transparent solution, free of sediment	Colorless and transparent solution, free of sediment
Odor(R151034D)	Irritating odor	Irritating Odor
Density(R151034D)(1.08 - 1.15 g/mL)	Conform	Conform
Appearance(R151034E)	Yellow and transparent solution, free of sediment	Yellow and transparent solution, free of sediment
Odor(R151034E)	Odorless	Odorless
Density(R151034E)(1.01 - 1.10 g/mL)	Conform	Conform
Appearance(R151034F)	Colorless and transparent solution, free of sediment	Colorless and transparent solution, free of sediment
Odor(R151034F)	Odorless	Odorless
Density(R151034F)(1.00 - 1.08 g/mL)	Conform	Conform
Odor(R151034B)	Odorless	Odorless



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

A handwritten signature in black ink that reads "Sophia Cheng".

Sophia Cheng
Quality Assurance