

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-03-28 08:17:14

Date of Next Testing: 2025-03-27 08:17:14

Product Name: (As,Bi,Cd,Cu,Cr,Fe,Hg,Mn,Ni,Se,Pb,Zn)multi-element mixed standard solution(mixed standard)

CAS Number:

Specifications & Purity: 100μg/ml solvent:5%Nitric acid,traceHCl

Lot #: C23241312

SKU # S117478

Version 0

Analysis #: 071519

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Concentration(each)	98.6-101.4 (mg/L)	100.00000 mg/L



John Su
QA & QC Manager