

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-01-04 11:39:47

Date of Next Testing: 2030-01-03 11:39:47

Product Name: (As,Sb,Bi,Ca,Co,Cu,Fe,Ga,K,Li,Mg,Mn,Na,Ni,Pb,Sr)multi-element mixed standard solution(mixed standard)

CAS Number:

Specifications & Purity: 100µg/ml, in hydrochloric acid, nitric acid

Lot #: L2526260

SKU #: S117480

Version: 1

Analysis #: 628149

Storage Temperature: Protected from light,Room temperature

Parameter	Limit Values	Result
As,Sb,Bi,Ca,Co,Cu,Fe,Ga,K,Li,Mg,Mn,Na,Ni,Pb,Sr	98.6-101.4 (mg/L)	100.00000 mg/L
Related expanded uncertainty(k=2)	0-4 (%)	1.40000 %
Solvent(15% HCl+5% HNO3)	Conform	Conform



John Su
QA & QC Manager