


ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-01-10 16:17:37
Date of Next Testing: 2024-01-10 16:17:37
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 101.0 mOsmol·Kg-1
Lot #: A2206502
SKU #: **S119603**
Analysis: **055343**

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	101.10 mOsmol/kg



John Su
QA & QC Manager