

Certificate of Analysis

Date of Testing: 2023-04-08 13:49:10
Date of Next Testing: 2025-04-07 13:49:10
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 101.0 mOsmol·Kg-1
Lot #: C2327251
SKU #: **S119603**
Analysis: **161778**

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	101.30 mOsmol/kg



John Su
QA & QC Manager