

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-04-28 13:43:24

Date of Next Testing: 2028-04-27 13:43:24

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 101.0 mOsmol·Kg-1

Lot #: D2623574

SKU #: **S119603**

Version: **1**

Analysis #: **679980**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	100.40000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %



John Su
QA & QC Manager