

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-05-27 17:10:29

Date of Next Testing: 2027-05-27 17:10:29

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 101.0 mOsmol·Kg-1

Lot #: E2522418

SKU #: **S119603**

Version: **1**

Analysis #: **514873**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	100.10000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %



John Su
QA & QC Manager