

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-07-29 17:12:08
Date of Next Testing: 2024-07-28 17:12:08
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 101.0 mOsmol·Kg-1
Lot #: G2228127
SKU #: **S119603**
Analysis: **105202**

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	100.20 mOsmol/kg



John Su
QA & QC Manager