

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-03-16 10:43:53
Date of Next Testing: 2025-03-15 10:43:53
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 101.0 mOsmol·Kg-1
Lot #: I1506150
SKU #: **S119603**
Analysis: **200134**

| Parameter | Limit Values | Result |
|------------|--------------------|------------------|
| Osmolality | 98-102 (mOsmol/kg) | 100.00 mOsmol/kg |

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager