

Certificate of Analysis

Date of Testing: 2024-10-22 14:10:21
Date of Next Testing: 2026-10-22 14:10:21
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 101.0 mOsmol·Kg-1
Lot #: J2411963
SKU #: **S119603**
Version: **1**
Analysis #: **412122**

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	100.00 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %



John Su
QA & QC Manager