

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-11-11 18:23:44

Date of Next Testing: 2027-11-11 18:23:44

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 101.0 mOsmol·Kg-1

Lot #: J2531640

SKU #: **S119603**

Version: **1**

Analysis #: **594717**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	98-102 (mOsmol/kg)	100.70000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %



John Su
QA & QC Manager