


ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-02-23 15:41:59
Date of Next Testing: 2024-02-23 15:41:59
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 199.9 mOsmol·Kg-1
Lot #: B2218096
SKU #: **S119604**
Analysis: **067222**

Parameter	Limit Values	Result
Osmolality	198-202 (mOsmol/kg)	200.00 mOsmol/kg



John Su
QA & QC Manager