

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-06-27 08:10:51
Date of Next Testing: 2028-06-26 08:10:51
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 199.9 mOsmol·Kg-1
Lot #: F2625408
SKU #: **S119604**
Version: **1**
Analysis #: **712936**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	198-202 (mOsmol/kg)	199.80000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %



John Su
QA & QC Manager