



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-09-23 19:21:54

Date of Next Testing: 2027-09-23 19:21:54

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 199.9 mOsmol·Kg-1

Lot #: I2515478

SKU # **S119604**

Version **1**

Analysis #: **546682**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	198-202 (mOsmol/kg)	200.00000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %

John Su
QA & QC Manager