

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-10-13 10:01:49

Date of Next Testing: 2025-10-12 10:01:49

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 199.9 mOsmol·Kg-1

Lot #: J2310380

SKU #: **S119604**

Analysis: 280930

Parameter	Limit Values	Result
Osmolality	198-202 (mOsmol/kg)	199.90 mOsmol/kg



John Su
QA & QC Manager