



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-01-04 14:00:54

Date of Next Testing: 2027-01-04 14:00:54

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 199.9 mOsmol·Kg-1

Lot #: L2420766

SKU #: **S119604**

Version: **1**

Analysis #: **444614**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	198-202 (mOsmol/kg)	200.40000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.50000 %

John Su
QA & QC Manager