

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-05-09 14:34:09
Date of Next Testing: 2025-05-08 14:34:09
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 299.5 mOsmol·Kg-1
Lot #: G2117287
SKU #: **S119605**
Analysis: 220193
Storage Temperature: Protected from light, Room temperature

Parameter	Limit Values	Result
Osmolality	298-302 (mOsmol/kg)	299.30 mOsmol/kg

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager