



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-04-17 10:56:31

Date of Next Testing: 2027-04-17 10:56:31

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 399.6 mOsmol·Kg-1

Lot #: D2508682

SKU #: **S119606**

Version: **1**

Analysis #: **489660**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	399.40000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.90000 %

John Su
QA & QC Manager