

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-04-29 08:49:01
Date of Next Testing: 2027-04-29 08:49:01
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 399.6 mOsmol·Kg-1
Lot #: D2523244
SKU #: **S119606**
Version: **1**
Analysis #: **498487**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	399.40 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	1.90000 %



John Su
QA & QC Manager