

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-04-15 15:16:53
Date of Next Testing: 2026-04-15 15:16:53
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 399.6 mOsmol·Kg-1
Lot #: G2207309
SKU #: **S119606**
Analysis: **341955**

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	400.17 mOsmol/kg



John Su
QA & QC Manager