

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2024-07-03 15:18:33  
Date of Next Testing: 2026-07-03 15:18:33  
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**  
CAS Number: 7647-14-5  
Specifications & Purity: 399.6 mOsmol·Kg-1  
Lot #: J22111354  
SKU #: **S119606**  
Version: **0**  
Analysis #: **379935**

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	400.07 mOsmol/kg

A handwritten signature in black ink that reads "John Su".

John Su  
QA & QC Manager