

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-10-30 16:13:43
Date of Next Testing: 2026-10-30 16:13:43
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 399.6 mOsmol·Kg-1
Lot #: J2425892
SKU #: **S119606**
Version: **1**
Analysis #: **417565**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	401.00 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	2.00000 %



John Su
QA & QC Manager