

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-06-18 14:30:12
Date of Next Testing: 2026-06-18 14:30:12
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 399.6 mOsmol·Kg-1
Lot #: L2007234
SKU #: **S119606**
Version: **0**
Analysis #: **371010**

Parameter	Limit Values	Result
Osmolality	398-402 (mOsmol/kg)	399.00 mOsmol/kg



John Su
QA & QC Manager