


ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-04-08 11:33:39
Date of Next Testing: 2026-04-08 11:33:39
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 500.5 mOsmol·Kg-1
Lot #: F2230399
SKU #: **S119607**
Analysis: **340030**

Parameter	Limit Values	Result
Osmolality	490-510 (mOsmol/kg)	507.88 mOsmol/kg



John Su
QA & QC Manager