

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-03-03 11:52:09

Date of Next Testing: 2027-03-03 11:52:09

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 600.9 mOsmol·Kg-1

Lot #: B2528515

SKU #: **S119608**

Version: **1**

Analysis #: **415575**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Osmolality	597-603 (mOsmol/kg)	600.90000 mOsmol/kg
Related expanded uncertainty(k=2)	0-2.8 (%)	2.80000 %



John Su
QA & QC Manager