

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2024-06-19 09:14:35  
Date of Next Testing: 2026-06-19 09:14:35  
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**  
CAS Number: 7647-14-5  
Specifications & Purity: 600.9 mOsmol·Kg-1  
Lot #: F2405500  
SKU #: **S119608**  
Version: **0**  
Analysis #: **334632**

Parameter	Limit Values	Result
Osmolality	597-603 (mOsmol/kg)	600.00 mOsmol/kg

A handwritten signature in black ink that reads "John Su".

John Su  
QA & QC Manager