

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-06-19 09:14:54
Date of Next Testing: 2026-06-19 09:14:54
Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**
CAS Number: 7647-14-5
Specifications & Purity: 600.9 mOsmol·Kg-1
Lot #: F2405501
SKU #: **S119608**
Version: **0**
Analysis #: **334632**

Parameter	Limit Values	Result
Osmolality	597-603 (mOsmol/kg)	600.00 mOsmol/kg



John Su
QA & QC Manager