



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-09-08 19:12:45

Date of Next Testing: 2027-09-08 19:12:45

Product Name: **Sodium Chloride Standard Solutions for Osmometer Calibration**

CAS Number: 7647-14-5

Specifications & Purity: 700.2 mOsmol·Kg-1

Lot #: I2502565

SKU # **S119609**

Version **1**

Analysis #: **558321**

Storage Temperature: Protected from light, Room temperature

Parameter	Limit Values	Result
Osmolality	697-703 (mOsmol/kg)	700.00000 mOsmol/kg
Related expanded uncertainty(k=2)	0-4 (%)	3.20000 %

John Su
QA & QC Manager