

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-06-23 14:58:28
Date of Next Testing: 2027-06-13 14:58:28
Product Name: **Silver standard for water analysis**
CAS Number: 7440-22-4(HNO3)
Specifications & Purity: analytical standard, 0.3 mg/L in 1%HNO3
Lot #: E1609025
SKU #: **S140982**
Version: **0**
Analysis #: **529651**

Parameter	Limit Values	Result
Concentration	0.29-0.31 (mg/L)	0.29 mg/L

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager