

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-04-09 17:15:46

Date of Next Testing: 2027-04-09 17:15:46

Product Name: **Spore Staining Solution (Moeller's Method)**

CAS Number:

Specifications & Purity: BioReagent, Biological Stain, Suitable for Microbiology, for microscopy

Lot #: ZJ26F0433899

SKU #: **S1510446-CG3×20ml**

Version: **1**

Analysis #: 030362

Parameter	Limit Value	Result
Appearance (S1510446A)	Red opaque liquid with a characteristic odor, no sediment	Red opaque liquid with a characteristic odor, no sediment
Density (S1510446A): 0.94~1.01(g/mL)	Conform	Conform
Appearance (S1510446B)	Colorless, transparent liquid with a characteristic odor and no sediment	Colorless, transparent liquid with a characteristic odor and no sediment
Density (S1510446B): 0.79~0.87(g/mL)	Conform	Conform
Appearance (S1510446C)	Blue-violet opaque liquid with a characteristic odor, no sediment or a small amount of sediment	Blue-violet opaque liquid with a characteristic odor, no sediment or a small amount of sediment
Density (S1510446C): 0.94~1.01(g/mL)	Conform	Conform
Seal Integrity Testing	Conform	Conform



Sophia Cheng
Quality Assurance