

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-05-11 13:38:11

Date of Next Testing: 2027-05-11 13:38:11

Product Name: **Picro Sirius Red Stain Kit**

CAS Number:

Specifications & Purity: BioReagent, Biological Stain, for microscopy

Lot #: ZJ26F0535212

SKU #: **S774173**

Version: **1**

Analysis #: 034525

Parameter	Limit Value	Result
Absorbance test (A502 nm after 1:20 dilution)	0.45-0.65	0.48370
Appearance(S774173)	Clear liquid	Clear liquid
Collagen staining	Conform	Conform



Sophia Cheng
Quality Assurance