



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-04-03 10:08:45
Date of Next Testing: 2028-04-02 10:08:45
Product Name: **Tungsten Flux**
CAS Number:
Specifications & Purity: Granular,for infrared carbon sulfur analyzer
Lot #: H1630021
SKU #: **T130834**
Version: **1**
Analysis #: **339223**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Total Metallic Impuritie	0-500 (ppm)	429.00000 ppm
Particle size	20-40 (mesh)	25.00000 mesh
S	0-0.0005 (%)	0.00009 %
Carbon(C)	0-0.0008 (%)	0.00009 %
Purity(Based On Trace Metals Analysis)	99.95-100 (%)	99.95000 %
X-Ray Diffraction	Conforms to Structure	Conforms to Structure
Applied to element analyzers	conforms	Conforms

John Su
QA & QC Manager