



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-08-19 14:44:42
Date of Next Testing: 2029-08-18 14:44:42
Product Name: **Tungsten Flux**
CAS Number:
Specifications & Purity: Granular,for infrared carbon sulfur analyzer
Lot #: K2104128
SKU #: **T130834**
Version **1**
Analysis #: **556996**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Total Metallic Impuritie	0-500 (ppm)	300.00000 ppm
Particle size	20-40 (mesh)	24.90000 mesh
S	0-0.0005 (%)	0.00020 %
Carbon(C)	0-0.0008 (%)	0.00034 %
Purity(Based On Trace Metals Analysis)	99.95-100 (%)	99.97000 %
X-Ray Diffraction	Conforms to Structure	Conforms to Structure
Applied to element analyzers	conforms	Conforms

John Su
QA & QC Manager