



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-01-10 13:40:06

Date of Next Testing: 2029-01-09 13:40:06

Product Name: **Technical pH Buffer solutions@25°C (pH 9.21±0.02)**

CAS Number:

Specifications & Purity: Technical Buffer Solution pH 9.21±0.02 @ 25°C, (Blue)

Lot #: A2510314

SKU # **T196896**

Version **0**

Analysis #: **461066**

Parameter	Limit Values	Result
PH(25 °C)	9.19-9.23	9.21
Appearance (T196896)	Blue transparency liquid	Blue transparency liquid

John Su
QA & QC Manager